

May 24, 1973

Dr. Frank W. Jackson
Cowley Associates
1919 North Front Street
Harrisburg, Pennsylvania 17102

Dear Frank:

You did a fine job with your exercise testing handbook.

I hope the next version will provide:

1. The "probabilistic" concept to interpretation
2. A set of simple measurements
3. A proposed classification of ST response

I was pleased that, if you had to use abnormal-normal terminology, that you defined them well.

I was mildly disappointed that you chose the 60 msec interval rather than the 0.5 ~~OR~~ amplitude measurement which is both independent of heart rate and more repeatable.

Cordially,

Henry Blackburn, M.D.
Professor and Director

cao